

TECHNICAL DATA

Fluke AC175 Alligator Clip Set



Key features

Insulated alligator clips with flexible boots

Threaded shaft to twist onto test probes

Jaw opening: 7 mm (0.275 in)

Jaw contact material: nickel-plated steel

Product overview: Fluke AC175 Alligator Clip Set

Modular clips are used with modular test leads to offer additional flexibility to configure your leads to your individual needs. The insulated alligator clip jaws provide a 7 mm opening with flexible boots.

The Fluke AC175 Threaded Alligator Clip Set comes with a one-year warranty and a cat rating of CAT III 1000 V, CAT IV 600 V, 10 A.

This alligator clip set has slide-on alligator clips (red, black) compatible with multiple Fluke test probes:

Fluke TL71 Premium Test Lead Set

Fluke TL75 Hard Point™ Test Lead Set

Fluke TL175 TwistGuard™ Test Leads

Specifications: Fluke AC175 Alligator Clip Set

Jaw opening	7 mm (0.275 in), jaw contact material: nickel-plated steel
Category Rating	CAT III 1000 V, CAT IV 600 V, 10 A
Operating temperature	-20°C to 55°C (-4°F to 131°F)
Environmental Information	Pollution Degree 2, 2000 m altitude

Ordering information



Fluke AC175

AC175 Alligator Clip Set

Optional accessories	Description
Fluke TL71 Premium Test Lead Set	Comfort grip, silicone insulated. μ V measurement recommended.
Fluke TL175 TwistGuard™ Test Leads	Simply twist to change the exposed probe tip length.
Fluke TP175 TwistGuard™ Test Probes	The Fluke TP175 TwistGuard® Test Probes have adjustable tip lengths that allow you to work safely in different settings.
Fluke TL75 Hard Point™ Test Lead Set	PVC-insulated, right-angle shrouded 4 mm (0.16 in) banana plugs

Fluke. *Keeping your world up and running.®*

Fluke (UK) Ltd.
52 Hurricane Way
Norwich, Norfolk
NR6 6JB
United Kingdom
Tel.: +44 (0)20 7942 0708
E-mail: cs.uk@fluke.com
www.fluke.com

©2021 Fluke Corporation. All rights reserved.
Data subject to alteration without notice.
11/2021

**Modification of this document is not permitted
without written permission from Fluke Corporation.**