

# Fluke i1000s AC Current Probe



## Key features

- Allows accurate measurement of currents from 100 mA to 1000 A rms, 5 Hz to 100 kHz without breaking into the circuit
- A passive filter eliminates noise and ring on rapidly rising di/dt waveform, ensuring accurate screen displays
- Connects directly to an oscilloscope through a reinforced coaxial cable and an insulated BNC connector
- Can be used with multimeters with optional PM9081/001 BNC/Banana adapter
- One year warranty

## Product overview: Fluke i1000s AC Current Probe

The Fluke i1000s is a clamp-on AC current probe designed to expand oscilloscope applications in industrial and power environments. Ideal for measuring distorted current waveforms and harmonics.

## Specifications: Fluke i1000s AC Current Probe

Specifications	
Nominal current range	10 A 100 A 1000 A

Continuous current range	0.1 A - 10 A 0.1 A - 100 A 1 A - 1000 A
Maximum non-destructive current	2000 A
Lowest measureable current	0.1
Basic accuracy	1% + 1 A (48-65 Hz) (% reading + floorspec)
Useable frequency	5 Hz - 100 kHz
Output level(s)	100 mv/A 10 mv/A 1 mv/A
<b>Safety Specifications</b>	
Safety	CAT III, 600 V CAT IV, 300V
Maximum voltage	600 V AC
<b>Mechanical and General Specifications</b>	
Warranty	1 year
Maximum conductor diameter	54 mm
Output cable length	1.6 m
BNC adapter	Yes

## Ordering information



### Fluke i1000s

---

**Fluke.** *Keeping your world up and running.®*

**Fluke Corporation**

PO Box 9090, Everett, WA 98206 U.S.A.

**Fluke Electronics (Malaysia) Sdn Bhd**

Suite D-3A-4, Block D  
SouthGage Commercial Centre, No2  
Jalan Dua Off Jln Chan Sow Lin  
55200 KL  
Malaysia

Email: [info.asean@fluke.com](mailto:info.asean@fluke.com)

[www.fluke.com/my](http://www.fluke.com/my)

©2021 Fluke Corporation. Specifications subject to  
change without notice.

12/2021

**For more information call:**

In the U.S.A. (800) 443-5853

In Europe/M-East/Africa +31 (0)40 267 5100

In Canada (800)-36-FLUKE

From other countries +1 (425) 446-5500

**Modification of this document is not permitted  
without written permission from Fluke Corporation.**